



M e a s u r e m e n t s L t d

Certificate of Calibration

Customer: Vaf- Fluid Technik GmbH
Gottfried Schenker Str 12
09244
Lichtenau
Germany

Certificate No S16176

:

Serial No. 14114/2

Flowmeter Type: Uniflux Glass Tube Variable Area

Flow Range: 2-25L/min O2

The above flowmeter was calibrated by series comparison
against master traceable to National Standards.
Master serial no. IMT102

Indicated Flow
O2 L/min
@20°C,
1.013 bar a

Measured Flow
O2 L/min
@20°C,
1.013 bar a

25	25.25
20	20.18
15	15.99
10	10.72
2	1.85

The best measurement capability expressed as an uncertainty does not exceed 1%.

The uncertainties are for a confidence probability of not less than 95%.

Signed: J Fox

Date: 6/7/2011

For Influx Measurements Ltd